

A member of Top Glove Group: The World's Largest Manufacturer of Gloves

FACTORY 36 : 2, Jalan 8, Pengkalan Chepa 2 Industrial Zone, 16100 Kota Bharu, Kelantan D.N., Malaysia.
☎ +609 774 7171 ☎ +609 771 3565/3072 ☎ +6012 2896 270 ✉ sales@topglove.com.my 🌐 www.topglove.com

EU DECLARATION OF CONFORMITY

Manufacturing Site : Terang Nusa (Malaysia) Sdn. Bhd.
: 2, Jalan 8, Pengkalan Chepa 2 Industrial Zone,
16100 Kota Bharu, Kelantan D.N., Malaysia.

European Authorized Representative : Ulma International GmbH
Pfaffenweg 35,
89231 Neu-Ulm, Germany.

Name of Device : Sterile Latex Surgical Powdered Gloves
Device Reference Code (PPER) : TNMSA101
Customer Reference Code : DONA SENSITIVE (9021)
Size : 5.5, 6.0, 6.5, 7.0, 7.5, 8.0, 8.5, 9.0
Classification (MDD) : Class IIa
Classification (PPER) : Category III
Conformity Assessment Procedure (MDD) : Annex II excluding (4)
Conformity Assessment Procedure (PPER) : Annex VII (Module C2)
EU Type Examination Certificate Number (PPER) : 2777/10576-03/E00-00
EU Type Examination Certificate Issued by (PPER) : SATRA Technology Europe Limited,
Bracetown Business Park,
Clonee, D15YN2P, Ireland.

Notified Body Number (PPER) : 2777
Notified Body Number (MDD) : 0123
Notified Body (MDD) : TÜV SÜD Product Service GmbH,
Ridlerstraße 65, 80339, Munich, Germany.

EC Certificate Number (MDD) : G1 061155 0014 Rev. 02

We Terang Nusa (Malaysia) Sdn Bhd herewith declare with our own responsibility that above mentioned product with CE mark;

- i. Meet the provisions of the EC Council Directive 93/42/EEC 14th June 1993 concerned medical devices, amended by Council Directive 2007/47/EC. All supporting documentations are retained under the premise of manufacturer.
- ii. is following to the EU Type Examination and conformity with the provisions of the new PPE Regulations (EU) 2016/425 Category III and, where such is the case, with the national standard transposing harmonized standard no. EN ISO 21420:2020, EN ISO 374-1:2016+A1:2018 (EN ISO 374-2:2019, EN 16523-1:2015+A1:2018, EN ISO 374-4:2019) and EN ISO 374-5:2016.
- iii. is subject to the procedures set out in Annex VII (Module C2) of the new PPE Regulations (EU) 2016/425 under the supervision of the notified body SATRA Technology Europe Limited, Bracetown Business Park, Clonee, D15YN2P, Ireland.

**“TO PREVENT CORRUPTION & BRIBERY. CORRUPTION & BRIBERY IS A CRIME.
BE HONEST AND NO CHEATING”**

DP 21/01/20/TGT

Applicable Standards:

No	Standard	Descriptions	Date Published
1.	EN ISO 13485:2016	Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2016)	March 2016
2.	EN 455-1:2020+1:2022	Medical gloves for single use. Part 1: Requirement and testing for freedom from holes.	February 2022
3.	EN 455-2:2015	Medical gloves for single use. Part 2: Requirement and testing for physical properties.	April 2015
4.	EN 455-3:2015	Medical gloves for single use. Part 3: Requirement and testing for biological evaluation.	April 2015
5.	EN 455-4:2009	Medical gloves for single use. Part 4: Requirements and testing for shelf life determination.	October 2009
6.	EN ISO14971:2019+A11:2021	Medical devices - Application of risk management to medical devices (ISO 14971:2019)	December 2021
7.	ISO 2859-1:2011	Sampling procedures and table for inspection by attributes	June 2011
8.	EN ISO 11737-1:2018/ A1:2021	Sterilization of health care products Microbiological methods - Part 1: Determination of a population of microorganisms on products - Amendment 1 (ISO 11737-1:2018/Amd 1:2021)	June 2021
9.	EN ISO 11737-2:2020	Sterilization of health care products - Microbiological methods - Part 2: Tests of sterility performed in the definition, validation and maintenance of a sterilization process (ISO 11737-2:2019)	May 2020
10.	EN ISO 11137-1:2015/ A2:2019	Sterilization of health care products - Radiation - Part 1: Requirements for development, validation and routine control of a sterilization process for medical devices - Amendment 2: Revision to 4.3.4 and 11.2 (ISO 11137-1:2006/Amd 2:2018)	November 2019
11.	EN ISO 11137-2:2015	Sterilization of health care products - Radiation - Part 2: Establishing the sterilization dose (ISO 11137-2:2013)	June 2015
12.	EN ISO 10993-1:2020	Biological evaluation for medical device – Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2018)	December 2020
13.	EN ISO 10993-5:2009	Biological evaluation of medical devices - Part 5: Tests for in vitro cytotoxicity (ISO 10993-5:2009)	June 2009
14.	EN ISO 10993-10:2013	Biological evaluation of medical devices - Part 10: Tests for irritation and skin sensitization (ISO 10993-10:2010)	August 2013
15.	EN ISO 10993-11:2018	Biological evaluation of medical devices. Test for systemic toxicity (ISO 10993-11:2017)	May 2018
16.	EN ISO 10993-23:2021	Biological evaluation of medical devices - Part 23: Tests for irritation (ISO 10993-23:2021)	March 2021
17.	EN ISO 11607-1:2020	Packaging for terminally sterilized medical devices - Part 1: Requirements for materials, sterile barrier systems and packaging systems (ISO 11607-1:2019)	January 2020
18.	EN ISO 11607-2:2020	Packaging for terminally sterilized medical devices - Part 2: Validation requirements for forming, sealing and assembly processes (ISO 11607-2:2019)	January 2020

19.	EN ISO 15223-1:2021	Medical devices - Symbols to be used with information to be supplied by the manufacturer - Part 1: General requirements (ISO 15223-1:2021)	September 2021
20.	ISO/TR 20416:2020	Medical devices — Post-market surveillance for manufacturers	July 2020
21.	MEDDEV 2.4/1	2.4/1 Classification of Medical Device	Revision 9, June 2010
22.	MEDDEV 2.5/10	2.5/10 Guideline for Authorized Representative	January 2012
23.	MEDDEV 2.7/1	2.7/1 Clinical Evaluation	Revision 4, June 2016
24.	MEDDEV 2.12/1	2.12/1 Medical Device Vigilance System	Revision 8, January 2013

EU DOC Issuance Date

: 27th May 2024



Name: Pn Noor Akilah Saidin
Designation: General Manager, RA

