



Medical Devices (Pvt) Ltd.

Wazirabad Road, Ugoki, Sialkot - Pakistan.

MD-QA-22-113

7th Nov 2022

Declaration of Conformity

We Medical Devices (Pvt) Ltd declares, under our sole responsibility, that the LARYNGOSCOPE which consists of the following Models and Types are of Class Is, and meet all the applicable provisions set out in Directive 93/42/EEC according to Annex V.

Article Description	Trade Name	GIMA Ref #	MD Catalog #
Laryngoscope Blade (Single Use), F.O. Miller No. 0	Lunalite 'D'	58001	810-411-00S
Laryngoscope Blade (Single Use), F.O. Miller No. 1	Lunalite 'D'	58002	810-411-01S
Laryngoscope Blade (Single Use), F.O. Macintosh No. 1	Lunalite 'D'	58006	810-401-01S
Laryngoscope Blade (Single Use), F.O. Macintosh No. 2	Lunalite 'D'	58007	810-401-02S
Laryngoscope Blade (Single Use), F.O. Macintosh No. 3	Lunalite 'D'	58008	810-401-03S
Laryngoscope Blade (Single Use), F.O. Macintosh No. 4	Lunalite 'D'	58009	810-401-04S
Laryngoscope Blade (Single Use), F.O. Macintosh No. 5	Lunalite 'D'	58010	810-401-05S
Laryngoscope Handle (Single Use), F.O. Small, V-2 "Ready to Use" with LED illumination.	LunaLED	58016	810-459-25S
Laryngoscope Handle (Single Use), F.O. Medium, V-2 "Ready to Use" with LED illumination.	LunaLED	58017	810-457-25S

Applied Standards: ISO 13485:2016, ISO 14971:2019, ISO 15223-1:2021 (E), ISO 10993-5: 2009, ISO 10993-10: 2010, ISO 7376:2020, ISO 14644-1:2015, IEC 60601-1-2:2014, ISO 10993-7:2008, ISO 11607-2:2019, ISO 11135:2014, CE 1639, IEC 60601-1-6, IEC 60601-1:2022, IEC 62366-1:2015.

EC Certificate Number: PK19/818842597.01, Valid Till 24th May 2024.


ISO 13485:2016 Certificate Number: PK04/61153.01, Valid Till 20th April 2025.

EU Representative: Obelis
S.A Bd. General Wahis 53
B-1030 Brussels, Belgium
Ph # 32.2.732.59.54

Notified Body: SGS Belgium NV;
SGS House Noorderlaan 87 2030
Antwerp Belgium
Tel: +32 (0)3 545-48-48

SIALKOT-PAKISTAN
Place (City & Country)

QUALITY ASSURANCE MANAGER
Position of Authorized person


Signature